



North Carolina State Laboratory Public Health
Environmental Sciences
Microbiology
Certificate of Analysis

P.O. Box 28047
306 N. Wilmington St.
Raleigh, NC 27611-8047
<http://slph.ncpublichealth.com>
Phone: 919-733-7834
Fax: 919-733-8695

Report To: **HENDERSON CO ENVIRONMENTAL HEALTH** Name of System: **GREG HOLT**
1200 SPARTANBURG HWY - SUITE 100 **171 WILD LAUREL WAY**
HENDERSONVILLE, NC 28792 **HORSE SHOE, NC 28742**
EIN:566000307EH COURIER #: **06-92-01**

StarLiMS Sample ID: **ES011912-0090001** Collected: **01/18/2012** 09:16 **Curt Cline**
 Received: **01/19/2012** 08:01 **Joy Hayes**
ES Microbiology ID: **33424** Sample Source: **Well** Well Permit Number: **SM 9223B**
GPS Number: Sampling Point: **Kitchen**

Sample Description:

Comment:

Environmental Microbiology - Colilert Profile Method: **SM 9223B**

Test Name: **Colilert**

Analyte	Test Result	Analyst	Date
Total Coliform, Colilert	Absent	Darneice Lyons	01/20/2012
<i>E. coli</i> , Colilert	Absent	Darneice Lyons	01/20/2012

Report Date: **01/20/2012**

Reported By: **Susan Beasley**



Explanations of Coliform Analysis:

If coliform bacteria are **Absent**, the water is considered safe for drinking purpose. If coliform bacteria are **Present**, the water is considered unsafe for drinking purpose. Presence of *E. coli* (bacteria) generally indicates that the water has been contaminated with fecal material. It must be remembered that a water analysis refers only to the sample received and should not be regarded as a complete report on the water supply.