



North Carolina State Laboratory Public Health

Environmental Sciences

Microbiology

Certificate of Analysis

P.O. Box 28047
306 N. Wilmington St.
Raleigh, NC 27611-8047
<http://slph.state.nc.us>
Phone: 919-733-7834
Fax: 919-733-8695

Report To:**HENDERSON CO ENVIRONMENTAL HEALTH****1200 SPARTANBURG HWY - SUITE 100
HENDERSONVILLE, NC 28792****Name of System:****Ricky Murray****620 Slick Rock Road
HENDERSONVILLE, NC 28792**StarLIMS Sample ID: **ES070209-0040001**

Collected: 07/01/2009 08:38

Received: 07/02/2009 07:47

John Goolsby**Benjamin Saavedra**ES Microbiology ID: **6659**

GPS Number:

Sample Source: **Well**Sampling Point: **Kitchen**

Well Permit Number:

Sample Description:

Comment:

Environmental Microbiology - Colilert Profile**Method:** SM 9223B**Test Name: Colilert**

Analyte	Test Result	Analyst	Date
Total Coliform, Colilert	Present	Benjamin Saavedra	07/06/2009
E. coli, Colilert	Present	Benjamin Saavedra	07/06/2009

Report Date: 07/06/2009**Reported By:** *Susan Beasley*