



North Carolina State Laboratory Public Health  
Environmental Sciences  
Microbiology  
Certificate of Analysis

P.O. Box 28047  
4312 District Drive  
Raleigh, NC 27611-8047  
<http://slph.ncpublichealth.com>  
Phone: 919-733-7308  
Fax: 919-715-8611

---

**Report To:** Name of System:  
**HENDERSON CO ENVIRONMENTAL HEALTH**  
**1200 SPARTANBURG HWY - STE 100** **FLETCHER VIEW INN**  
**HENDERSONVILLE, NC 28792** **51 FLETCHER VIEW DR**  
**EIN:566000307EH** **COURIER #: 06-92-01** **FLETCHER, NC 28732**

---

StarLiMS Sample ID: **ES071316-0063001** Collected: 07/12/2016 08:00 **S Foster**  
 Received: 07/13/2016 08:49 **Susan Beasley**

ES Microbiology ID: Sample Source: **Well** Well Permit Number:  
GPS Number: Sampling Point: **Kitchen tap**

Sample Description:

Comment:

---

**Environmental Microbiology - Colilert Profile** Method: SM 9223B

---

**Test Name:** Colilert

Analyte	Test Result	Analyst	Date
Total Coliform, Colilert	Absent	Denise Richardson	07/14/2016
<i>E. coli</i> , Colilert	Absent	Denise Richardson	07/14/2016

Report Date: 07/15/2016

Reported By: Susan Beasley

**Explanations of Coliform Analysis:**

If coliform bacteria are **Absent**, the water is considered safe for drinking purpose. If coliform bacteria are **Present**, the water is considered unsafe for drinking purpose. Presence of *E. coli* (bacteria) generally indicates that the water has been contaminated with fecal material. It must be remembered that a water analysis refers only to the sample received and should not be regarded as a complete report on the water supply.