



North Carolina State Laboratory of Public Health

Environmental Sciences

Microbiology

Certificate of Analysis

4312 District Drive
MSC 1918
Raleigh, NC 27699-1918

<http://slph.ncpublichealth.com>
Phone: 919-733-7308
Fax: 919-715-8611

FINAL REPORT

Report to: **CODY ROBERTSON**

Name of System:

HENDERSON CO ENVIRONMENTAL HEALTH
1200 SPARTANBURG HWY - STE 100
Hendersonville, NC 28792

David Coston
617 Apple Valley Rd

EIN: 566000307EH

Delivery: NC Courier

Henderson County

StarLiMS ID: **ES190516-0067** Date Collected: 05/15/2019 Time Collected: 09:00 By: Cody Robertson
Date Received: 05/16/2019 Time Received: 08:13 By: Tyler Best

Sample Source: Well water Sampling Point: Sink

Sample Type: GPS No.

Treatment: Well Permit No. 19090102959

Comment:

Colilert Profile

Method: SM 9223B

Analyte	Test Result	Unit	Conclusion	Date Tested
Total Coliform	Absent			05/16/2019
E. coli	Absent			05/16/2019

Report Date: 05/20/2019

Reported By: **Susan Beasley**

Explanations of Coliform Analysis:

If coliform bacteria are **Absent**, the water is considered safe for drinking purpose. If coliform bacteria are **Present**, the water is considered unsafe for drinking purpose. Presence of *E. coli* (bacteria) generally indicates that the water has been contaminated with fecal material. It must be remembered that a water analysis refers only to the sample received and should not be regarded as a complete report on the water supply.