



North Carolina State Laboratory of Public Health
Environmental Sciences
Microbiology
Certificate of Analysis

4312 District Drive
MSC 1918
Raleigh, NC 27699-1918
<http://slph.ncpublichealth.com>
Phone: 919-733-7308
Fax: 919-715-8611

FINAL REPORT

Report to: Nicki Bishop

Name of System:

HENDERSON CO ENVIRONMENTAL HEALTH
1200 SPARTANBURG HWY - STE 100
Hendersonville, NC 28792

Don Hill
1878 Old Clear Creek Rd

EIN: 566000307EH

Delivery: NC Courier

Henderson County

StarLiMS ID: ES190905-0071

Date Collected: 09/04/2019

Time Collected: 09:00

By:

Date Received: 09/05/2019

Time Received: 08:23

By: Angela Heybroek

Sample Source: New Well

Sampling Point:

Sample Type: Raw

GPS No.

Treatment:

Well Permit No. 18100111159

Comment:

Colilert Profile

Method: SM 9223B

Analyte	Test Result	Unit	Conclusion	Date Tested
Total Coliform	Present			09/05/2019
E. coli	Absent			09/05/2019

Report Date: 09/06/2019

Reported By: KPLEMMONS

Explanations of Coliform Analysis:

If coliform bacteria are **Absent**, the water is considered safe for drinking purpose. If coliform bacteria are **Present**, the water is considered unsafe for drinking purpose. Presence of *E. coli* (bacteria) generally indicates that the water has been contaminated with fecal material. It must be remembered that a water analysis refers only to the sample received and should not be regarded as a complete report on the water supply.