



North Carolina State Laboratory of Public Health

Environmental Sciences

Microbiology

Certificate of Analysis

4312 District Drive
MSC 1918
Raleigh, NC 27699-1918

<http://slph.ncpublichealth.com>
Phone: 919-733-7308
Fax: 919-715-8611

FINAL REPORT

Report to: Ronda Dufour

Name of System:

HENDERSON CO ENVIRONMENTAL HEALTH
1200 SPARTANBURG HWY - STE 100
Hendersonville, NC 28792

Roy Stepp
30 Ottanola Rd.
Hendersonville, NC 28792

EIN: 566000307EH

Delivery: NC Courier

Henderson County

StarLiMS ID: ES210616-0007	Date Collected: 06/15/2021	Time Collected: 08:50	By: Ronda Dufour
	Date Received: 06/16/2021	Time Received: 07:26	By: Denise Richardson
Sample Source: Well water	Sampling Point: Kitchen sink		
Sample Type:	GPS No.		
Treatment:	Well Permit No. 21090110475		

Comment:

Colilert Profile

Method: SM 9223B

Analyte	Test Result	Unit	Conclusion	Date Tested
Total Coliform	Not Tested			06/16/2021
E. coli	Not Tested			06/16/2021

Note: Analyte(s) 'Unsatisfactory' because length of time between sample collection and receipt exceeds acceptable requirements for this test..

Report Date: 06/17/2021

Reported By: **KPLEMMONS**

Explanations of Coliform Analysis:

If coliform bacteria are **Absent**, the water is considered safe for drinking purpose. If coliform bacteria are **Present**, the water is considered unsafe for drinking purpose. Presence of *E. coli* (bacteria) generally indicates that the water has been contaminated with fecal material. It must be remembered that a water analysis refers only to the sample received and should not be regarded as a complete report on the water supply.