



North Carolina State Laboratory of Public Health

Environmental Sciences

Microbiology

Certificate of Analysis

4312 District Drive
MSC 1918
Raleigh, NC 27699-1918

<http://slph.ncpublichealth.com>
Phone: 919-733-7308
Fax: 919-715-8611

FINAL REPORT

Report to: Jeffery Anders

Name of System:

HENDERSON CO ENVIRONMENTAL HEALTH
1200 SPARTANBURG HWY - STE 100
Hendersonville, NC 28792

D&D Gilbert Orchards
238 and 256 Howard Gilbert Ln
Hendersonville, NC 28792

EIN: 566000307EH

Delivery: NC Courier

Henderson County

StarLiMS ID: **ES230525-0035**

Date Collected: 05/23/2023

Time Collected: 09:00

By: J. Andres

Date Received: 05/25/2023

Time Received: 08:04

By: Shannon McCorkle

Sample Source: Well water

Sampling Point: House

Sample Type:

GPS No.

Treatment:

Well Permit No. 055-2023-0098/99

Comment:

Colilert Profile

Method: SM 9223B

Analyte	Test Result	Unit	Conclusion	Date Tested
Total Coliform	Not Tested			05/25/2023
E. coli	Not Tested			05/25/2023

Note: Analyte(s) 'Unsatisfactory' because exceeded EPA allowable holding time. Sample over 30 hrs old when received. Please resample..

Report Date: 05/26/2023

Reported By: **KPLEMMONS**

Explanations of Coliform Analysis:

If coliform bacteria are **Absent**, the water is considered safe for drinking purpose. If coliform bacteria are **Present**, the water is considered unsafe for drinking purpose. Presence of *E. coli* (bacteria) generally indicates that the water has been contaminated with fecal material. It must be remembered that a water analysis refers only to the sample received and should not be regarded as a complete report on the water supply.