



North Carolina State Laboratory of Public Health
Environmental Sciences
Microbiology
Certificate of Analysis

4312 District Drive
 MSC 1918
 Raleigh, NC 27699-1918
 http://slph.ncpublichealth.com
 Phone: 919-733-7308
 Fax: 919-715-8611

FINAL REPORT

Report to: Environmental Health

Name of System:

HENDERSON CO ENVIRONMENTAL HEALTH
 1200 SPARTANBURG HWY - STE 100
 Hendersonville, NC 28792

Jerry R Bishop
 5277 Big Hungry Rd
 Flat Rock, NC 28731

EIN: 566000307EH

Delivery: NC Courier

Henderson County

StarLiMS ID: ES241113-0079	Date Collected: 11/12/2024	Time Collected: 08:45	By: Jerry R Bishop
Barcode ID:	Date Received: 11/13/2024	Time Received: 08:15	By: Julie Schiavone
Sample Source: Hurricane Helene	Sampling Point: Spring		
Sample Type:	GPS No.		
Treatment: Chlorinated	Well Permit No.		
	Days Since Chlorine Treatment	>7 Days	

Comment:

Colilert - Quanti-Tray2000 Profile

Method: SM 9223B

Analyte	Test Result	Unit	Conclusion	Date Tested
Total Coliform	307.6	MPN/100mL		11/13/2024
E. coli	1	MPN/100mL		11/13/2024

Report Date: 11/14/2024

Reported By: **KKURTZ**

Explanations of Coliform Analysis:

If coliform bacteria are **Absent**, the water is considered safe for drinking purpose. If coliform bacteria are **Present**, the water is considered unsafe for drinking purpose. Presence of *E. coli* (bacteria) generally indicates that the water has been contaminated with fecal material. It must be remembered that a water analysis refers only to the sample received and should not be regarded as a complete report on the water supply.